

April 08, 2024

Attention: Bernie Hansen
ENGRX CORP
4929 74 AVENUE
EDMONTON, AB T6B 2H5

The design submission, Tracking Number 2024-01846, Web Portal Number 2024-S1278, originally received on March 26, 2024 was surveyed and accepted for registration as follows:

CRN : 0F23989.2 **Accepted on:** April 08, 2024

Reg Type: NEW DESIGN **Expiry Date:** April 08, 2034

Drawing No. : 645S, 745S, TM100, 645i, 745i, TM500 As Noted

Fitting type: Meter

Design registered in the name of : SIERRA INSTRUMENTS INC.

The registration is conditional on your compliance with the following notes:

The registration is based on identical design as 0F20194.2.

The code of construction is B31.3.

As indicated on AB-41 Statutory Declaration or AB-351 Declaration of Conformity form and submitted documentation, the code of construction is ASME B31.3.

- It is our understanding that the fitting(s), included as the scope of this submission, that is(are) subject to the Safety Codes Act shall comply with the requirements of the indicated Standard or Code of Construction on the AB-41 Statutory Declaration or AB-351 Declaration of Conformity as supported by the attached data which identifies the dimensions, materials of construction, press./temp. ratings and the basis for such ratings, and the identification marking of the fittings.

- This registration is valid only for fittings fabricated at the location(s) covered by the QC certificate attached to the accepted AB-41 Statutory Declaration or AB-351 Declaration of Conformity form.

- This registration is valid only until the indicated expiry date and only if the Manufacturer maintains a valid quality management system approved by an acceptable third-party agency, and maintains a valid Certification of Authorization Permit if required by the jurisdiction where manufacturing takes place, until that date.

- Should the approval of the quality management system lapse before the expiry date indicated above, this registration shall become void.

An invoice covering survey and registration fees will be forwarded from our Revenue Accounts.

If you have any question don't hesitate to contact me by phone at (780) 433-0281 ext 3388 or fax (780) 437-7787 or e-mail Liu@absa.ca.

Sincerely,



LIU, XING, P. Eng.
DOP Cert. No. D00008861