



Member of the FM Global Group

FM Approvals
1151 Boston Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781-762-9375 www.fmapprovals.com

CERTIFICATE OF COMPLIANCE

HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

240i-a-2-b-c-d-e-f-ST-g h-i. Inline Mass Vortex Flowmeter.

a = Multivariable V, VT or VTP.

b = Flow body F4, F5, F6, F7, F8, F9, F10, G4, G5, G6, G7, G8, G9, G10, H4, H5, H6, H7, H8, H9, H10, FD4, FD5, FD6, FD7, FD8, FD9, FD10, GD4, GD5, GD6, GD7, GD8, GD9, GD10, HD4, HD5, HD6, HD7, HD8, HD9 or HD10.

c = Enclosure E2 or E4(*). *Specify cable length (50 feet maximum).

d = Display DD or NR.

e = Power P2 or PS.

f = Output V6, V6H, V6M, V6DP2 or V6FF.

g = Process pressure MP0†, MP1*, MP2*, MP3* or MP4*. †V or VT only. *VTP only.

h = Certificates BLANK or MC, PT, CC or NC.

i = Oxygen cleaning BLANK or O2C.

Specific Condition of Use

For probes longer than 13 inches, a suitably certified conduit sealing device is required to be installed between the main enclosure and the probe body. The seal shall be located within 13 inches of the main enclosure.

241i-a-2-b-c-d-e-f-ST-g-h i-j-k. Insertion Mass Vortex Flowmeter.

a = Multivariable V, VT or VTP.

b = Probe length LS or LC.

c = Enclosure E2 or E4(*). *Specify cable length (50 feet maximum).

d = Display DD or NR.

e = Power P2 or PS.

f = Output V6, V6H, V6M, V6DP2 or V6FF.

g = Process pressure MP0†, MP1*, MP2*, MP3* or MP4*. †V or VT only. *VTP only.

h = Process connection CO, CM, CF, CG, CH, PM, PF, PG, PH, CFD, CGD, CHD, PFD, PGD or PHD.

i = Certificates BLANK or MC, PT, CC or NC.

j = Oxygen cleaning BLANK or O2C.

k = Retractor BLANK or 241-Retractor.

Specific Condition of Use

For probes longer than 13 inches, a suitably certified conduit sealing device is required to be installed between the main enclosure and the probe body. The seal shall be located within 13 inches of the main enclosure.



Member of the FM Global Group

Equipment Ratings:

Explosionproof for Class I, Division 1, Groups B, C and D hazardous (classified) locations, indoors and outdoors (Type 4X) with a temperature class rating of T6 for use in an ambient temperature range of -40°C to +60°C and process temperature range of -40°C to +200°C at a maximum pressure of 1500 psi.

FM Approved for:

Sierra Instruments Inc.
Monterey, CA, USA



Member of the FM Global Group

This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

Class 3600	2011
Class 3615	2006
Class 3810	2005
ANSI/NEMA 250	1991

Original Project ID: 3054986

Approval Granted: September 9, 2015

Subsequent Revision Reports / Date Approval Amended

Report Number	Date	Report Number	Date
---------------	------	---------------	------

FM Approvals LLC

J.E. Marquedant
Manager, Electrical Systems

9 September 2015

Date