

CERTIFICATE OF CONFORMITY



1. HAZARDOUS LOCATION ELECTRICAL EQUIPMENT PER CANADIAN REQUIREMENTS

2. Certificate No: FM18CA0034X

3. Equipment:
(Type Reference and Name) 640i and 780i Insertion and In-line Thermal Mass Flowmeters

4. Name of Listing Company: Sierra Instruments Inc

5. Address of Listing Company: 20 Ryan Ranch Road, Ste 109, Monterey, California 93940, United States of America

6. The examination and test results are recorded in confidential report number:

3048285C dated 25th April 2014

7. FM Approvals LLC, certifies that the equipment described has been found to comply with the following Approval standards and other documents:

CSA C22.2 No. 0.4:2017 (R2022), CSA C22.2 No. 0.5:2016 (R2020), CSA C22.2 No. 30:1986 (R2016),
CSA C22.2 No. 94:1991 (R2011), CSA C22.2 No. 61010-1:2012

8. If the sign 'X' is placed after the certificate number, it indicates that the equipment is subject to specific conditions of use specified in the schedule to this certificate.

9. This certificate relates to the design, examination and testing of the products specified herein. The FM Approvals surveillance audit program has further determined that the manufacturing processes and quality control procedures in place are satisfactory to manufacture the product as examined, tested and Approved.

10. Equipment Ratings:

Explosionproof for Class I, Division 1, Groups B, C and D hazardous locations, indoors and outdoors (Type 4X) for use in an ambient temperature range of -40°C to +60°C with temperature class marking of T3C.

11. The marking of the equipment shall include:

Class I Division 1, Groups B, C, D; T3C Ta = -40°C to +60°C; Type 4X

Certificate issued by:

J.E. Marquedant
VP, Manager - Electrical Systems

3 January 2026

Date

To verify the availability of the Approved product, please refer to www.approvalguide.com

THIS CERTIFICATE MAY ONLY BE REPRODUCED IN ITS ENTIRETY AND WITHOUT CHANGE

FM Approvals LLC. One Technology Way, Norwood, MA 02062 USA

T: +1 (1) 781 762 4300 F: +1 (1) 781 762 9375 E-mail: information@fmaprovals.com www.fmaprovals.com

F 348 (Jul 24)



SCHEDULE

Canadian Certificate of Conformity No: FM18CA0034X

12. Description of Equipment:

The 640i and 780i Insertion and In-line Thermal Mass Flowmeters are used for gas flow measurement applications. The flowmeters integrate configuration, validation and diagnostics in a probe-mounted or remotely-mounted housing. When remotely-mounted, the electronics enclosure can be mounted up to 200 feet away from the remote probe enclosure. All variables can be monitored on a viewable display. The flowmeters can also provide 4-20 mA flow and temperature output, alarm output and an optional pressure output. The 640i is suitable for insertion into pipes or ducts from three inches up to 72 inches while the 780i can go in-line to 8-inch flow bodies. For probes longer than 13 inches, a seal is required within 13 inches of the probe and enclosure.

The 640i and 780i Insertion and In-line Thermal Mass Flowmeters operate on 24 Vdc or 100-240 Vac. The flowmeters are intended for use in an ambient temperature of -40°C to +60°C and with process temperatures of -40°C to +200°C at a pressure of 300 psi (in-line) or 500 psi (insertion).

Refer to the Annex for model codes.

13. Specific Conditions of Use:

1. A suitably certified conduit sealing device is required within 13 inches of the enclosure for probes longer than 13 inches.
2. A suitably certified conduit sealing device is required within 18 inches of the main enclosure wiring compartment.

14. Test and Assessment Procedure and Conditions:

This Certificate has been issued in accordance with FM Approvals Canadian Certification Scheme.

15. Schedule Drawings

A copy of the technical documentation has been kept by FM Approvals.

16. Certificate History

Details of the supplements to this certificate are described below:

Date	Description
25 April 2014	Original Issue.
8 June 2018	<u>Supplement 4:</u> Report Reference: – PR449880 dated 8 June 2018 Description of the Change: Standards updates and certificate updated to the latest format.

To verify the availability of the Approved product, please refer to www.approvalguide.com

THIS CERTIFICATE MAY ONLY BE REPRODUCED IN ITS ENTIRETY AND WITHOUT CHANGE

FM Approvals LLC. One Technology Way, Norwood, MA 02062 USA

T: +1 (1) 781 762 4300 F: +1 (1) 781 762 9375 E-mail: information@fmaprovals.com www.fmaprovals.com

F 348 (Jul 24)



SCHEDULE

Canadian Certificate of Conformity No: FM18CA0034X

Date	Description
20 March 2024	<u>Supplement 5:</u> Report Reference: RR239863 dated 20 March 2024. Description of the Change(s): Revised model code related to software changes, updated non-safety related drawings
8 April 2025	<u>Supplement 6:</u> Report Reference: RR245487 dated 8 April 2025. Description of the Change(s): Updating company address. Other minor documentation changes not affecting the equipment safety. Minor model code corrections for errors and obsolescences. CSA No. 5 was entered in error on last supplement and has been replaced by CSA No. 0.5. Where possible, standards have been updated to the latest reaffirmed dates.
3 January 2026	<u>Supplement 7:</u> Report Reference: RR249007 dated 3 January 2026. Description of the Change(s): Model code reformatting/restructuring. No new features added that would affect the equipment safety.

To verify the availability of the Approved product, please refer to www.approvalguide.com

THIS CERTIFICATE MAY ONLY BE REPRODUCED IN ITS ENTIRETY AND WITHOUT CHANGE

FM Approvals LLC. One Technology Way, Norwood, MA 02062 USA

T: +1 (1) 781 762 4300 F: +1 (1) 781 762 9375 E-mail: information@fmapprovals.com www.fmapprovals.com

F 348 (Jul 24)



SCHEDULE

Canadian Certificate of Conformity No: FM18CA0034X

FM Approvals

ANNEX

640i Insertion Thermal Mass Flowmeter

Description of Equipment:

640i-a-b-c-d-DD-e-UC-00-f-f-g-h-i-j-k. Insertion Thermal Mass Flowmeter.

a = Multivariable: VT or VTP.

b = Probe: L06, L09, L13, L18, L24, L36, L48, L* or L*M5. *Specify probe length (max 48 inches).

c = Enclosure: E2 or E4.

d = Power: P2 or P3.

e = Digital communication: MB, BH or XX*. *None.

f = Gas: 00A*, 01*, 01A*, 02*, 02A*, 03*, 04*, 04A*, 06*, 06A*, 07*, 07A*, 08*, 08A*, 10*, 10A*, 11*, 12*, 12A*, 14*, 99*, 99A* or BLANK. *Add LF suffix to each gas code that has an ultra low flow calibration range.

g = Non-resettable totalizer option: NRT or BLANK.

h = qMix software license option: qMix or BLANK.

i = Remote cable option: CXXX*-VT, CXXX*-VTP or BLANK[†]. *Specify cable length in XXX feet. [†]E2 only.

j = Pressure MP1*, MP2*, MP3*, MP4* or BLANK[†]. *VTP only. [†]VT only.

k = Oxygen cleaning option: O2C or BLANK.

780i In-line Thermal Mass Flowmeter

Description of Equipment:

780i-a-b-c-d-DD-e-UC-00-f-f-g-h-i-j-k. In-line Thermal Mass Flowmeter.

a = Multivariable: VT or VTP.

b = Flow body: N050, N075, N100, N150, N200, N300, N400, N600, N800, F050, F075, F100, F150, F200, F300, F400, F600, F800, FD200, FD300, FD400, FD600, FD800, GD100, GD150, GD200 or GD300.

c = Enclosure: E2 or E4.

d = Power: P2 or P3.

e = Digital communication: MB, BH or XX*. *None.

f = Gas: 00A*, 01*, 01A*, 02*, 02A*, 03*, 04*, 04A*, 06*, 06A*, 07*, 07A*, 08*, 08A*, 10*, 10A*, 11*, 12*, 12A*, 14*, 99*, 99A* or BLANK. *Add LF suffix to each gas code that has an ultra low flow calibration range.

g = Non-resettable totalizer option: NRT or BLANK.

h = qMix software license option: qMix or BLANK.

i = Remote cable option: CXXX*-VT, CXXX*-VTP or BLANK[†]. *Specify cable length in XXX feet. [†]E2 only.

j = Pressure MP1*, MP2*, MP3*, MP4* or BLANK[†]. *VTP only. [†]VT only.

k = Oxygen cleaning option: O2C or BLANK.

To verify the availability of the Approved product, please refer to www.approvalguide.com

THIS CERTIFICATE MAY ONLY BE REPRODUCED IN ITS ENTIRETY AND WITHOUT CHANGE

FM Approvals LLC. One Technology Way, Norwood, MA 02062 USA

T: +1 (1) 781 762 4300 F: +1 (1) 781 762 9375 E-mail: information@fmaprovals.com www.fmaprovals.com

F 348 (Jul 24)

